

Calibration Device of Electrostatic Discharge Generator ESDD-A



In compliance with

- > IEC 61000-4-2
- > EN 61000-4-2
- > GB/T 17626.2

Introduction

The electrostatic discharge calibration test device is designed and developed to calibrate the electrostatic discharge generator according to the special needs of customers. The calibration device is made of solid wood structure platform, which conforms to IEC61000-4-2 and GB/T17626.2 standards. All parts are detachable, which is convenient for customers to move and test.

Parameters

Test desk	1700×1200×700 mm
Vertical Coupling Plate	1580×800×1590 mm
Electrostatic Gun Bracket	700×700×920 mm
Combined Size	3000×1580×1590 mm



SUZHOU 3CTEST ELECTRONIC CO., LTD.

Add.: No. 99 E'meishan Road, SND, Suzhou, Jiangsu Province, China

Tel: +86 (0)512 6807 7192 Fax: +86-512-68079795

Sales Email: globalsales@3ctest.cn Service Email: service@3ctest.cn

www.3c-test.com